



NATIONAL
ENVIRONMENTAL
TESTING, INC.

MAY 20 1998

Indianapolis Division
6964 Hillsdale Ct.
Indianapolis, IN 46250
Tel: (317) 842-4261
Fax: (317) 842-4286

ANALYTICAL REPORT

Mr. Richard Tyler
MILBANK MANUFACTURING INC
1400 E. HAVENS ST.
KOKOMO, IN 56901-3188

05/15/1998

NET Job Number: 98.03135

Page 1 of 3

Enclosed are the Analytical Results for the following samples submitted to NET, Inc. Indianapolis Division for analysis:

Project Description: TWICE WEEKLY WASTEWATER

Sample Number	Sample Description	Date Taken	Date Received
207039	OUTFALL 001	05/08/1998	05/08/1998

National Environmental Testing, Inc. certifies that the analytical results contained herein apply only to the specific samples analyzed.

Reproduction of this analytical report is permitted only in its entirety.

Project Representative

MIL0003223



NATIONAL
ENVIRONMENTAL
TESTING, INC.

MAY 20 1998

Indianapolis Division
6964 Hillsdale Ct.
Indianapolis, IN 46250
Tel: (317) 842-4261
Fax: (317) 842-4286

ANALYTICAL REPORT

Mr. Richard Tyler
MILBANK MANUFACTURING INC
1400 E. HAVENS ST.
KOKOMO, IN 56901-3188

05/15/1998

Job No.: 98.03135

Page 2 of 3

Date Received: 05/08/1998

Job Description: TWICE WEEKLY WASTEWATER

Sample Number / Sample I.D.	Sample Date/	Analyst &		Reporting		
Parameters	Result	Flag	Units	Date Analyzed	Method	Limit
207039	OUTFALL 001		05/08/1998			
Iron, ICP	0.24		mg/L	tjg / 05/12/1998	E-200.7	<0.10
Lead, ICP	<0.080		mg/L	tjg / 05/12/1998	E-200.7	<0.080
Silver, ICP	<0.020		mg/L	tjg / 05/12/1998	E-200.7	<0.020
Zinc, ICP	1.6		mg/L	tjg / 05/12/1998	E-200.7	<0.020

MIL0003224



NATIONAL
ENVIRONMENTAL
TESTING, INC.

MAY 20 1998

Indianapolis Division
6964 Hillsdale Ct.
Indianapolis, IN 46250
Tel: (317) 842-4261
Fax: (317) 842-4286

FIELD REPORT

JOB #: 98.03135
CLIENT: MILBANK MFG.
PROJECT: TWICE WEEKLY WASTEWATER
DATE: 5/08/98
SAMPLER(S): MTM

An ISCO model 6700 auto sampler was used in the composite mode of operation. The sampler was equipped with a plastic composite jug, tygon suction line, battery, ice, and strainer.

All reusable equipment is decontaminated withalconox, tap water, 5% nitric acid, and deionized water. New tygon suction tubing was used for the sampler. A stainless steel strainer was also used for the sampling event.

The sampler was set to take a sample every 30 minutes for 24 hours.

Monitoring start 08:00 on 5/07/98
Monitoring end 08:00 on 5/08/98

The composite sample was slightly cloudy.

Notes: Samples were preserved, in the field, in the appropriate containers with any required preservative.

MIL0003225



COMPANY MILBANK MFG.

PHONE (765) 452-5694 FAX _____

PROJECT NAME/LOCATION Twice Weekly Wastewater

PROJECT NUMBER

PROJECT MANAGER

REPORT TO: RICHARD TYLER

INVOICE TO: _____

P.O. NO. _____

NET QUOTE NO. 98.0060

SAMPLED BY

Michael Meyer
(PRINT NAME)

(PRINT NAME)

Michael Meyer
SIGNATURE

SIGNATURE

ANALYSES

To assist us in selecting the proper method

Is this work being conducted for regulatory compliance monitoring?

Yes _____ No _____

Is this work being conducted for regulatory enforcement action?

Yes _____ No _____

Which regulations apply: RCRA _____ NPDES Wastewater _____
UST _____ Drinking Water _____
Other _____ None _____

COMMENTS

MAY 20 1998

[illegible]

CONDITION OF SAMPLE: BOTTLES INTACT? YES / NO
FIELD FILTERED? YES / NO

COC SEALS PRESENT AND INTACT? ~~YES~~ / NO
VOLATILES FREE OF HEADSPACE? ~~YES~~ / NO

TEMPERATURE UPON RECEIPT: 2.5°C
Bottles supplied by NET? YES NO

SAMPLE REMAINDER DISPOSAL: RETURN SAMPLE REMAINDER TO CLIENT VIA _____
I REQUEST NET TO DISPOSE OF ALL SAMPLE REMAINDERS _____

DATE _____

RELINQUISHED BY:

DATE _____

TIME

RECEIVED BY:

RELINQUISHED BY:

DATE

	TIME
--	------

RECEIVED FOR NET BY:

METHOD OF SHIPMENT

REMARKS: